

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

KENNY SKIP HIRE
LESTER ROAD
NORSLEY MANOR ESTATE
LITTLE HULTON
MANCHESTER M38 0PT
LOCATION

Machine Details

Cust ref	
Make	BILANCIAI
Model	D410
Ser No	1130203731
Capacity	50/000 / kg
e/d	120 / kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
T11	9039	02/10/21

Cert Interval	ANNUALLY
Next Cert	FEB 2023

Tolerance +/-	kg
0-10000	20
10000-40000	40
40000-50000	60

Comments PLT - 5.5kg
WAGON - 22060

2	4	6
1	3	5
4		

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
2000	2000	0	2000	0
6000	5980	20	5980	20
10000	9980	20	9980	20
14000	13980	20	13980	20
18000	16000	0	16000	0
21560	21540	20	21540	20
22060	22060	0	22060	0
43620	43600	20	43600	20

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability 43620

Load	Ind	Ind	Ind
43600	43600	43600	

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 7560

Reading	Dev	Reading	Dev
1 7540	20	2 7560	0
3 7560	0	4 7560	0
5 7560	0	6 7560	0
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature [Signature] Name D. Pearce Date 10/02/2022
C1.3.88.1M