

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

KENNY WASTE MANAGEMENT
LESTER ROAD
WORSLEY TRADING EST
LITTLE HULTON
MANCHESTER
LOCATION M38 OPT
SITE

Machine Details

Cust ref	WEIGH62109F
Make	SBR-M
Model	D410
Ser No	T130203931
Capacity	50/000/lb
e/d	1/20/lb
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
T11-1-22	8478	8.10.20
T60-69	8479	8.10.20

Cert Interval	YEARLY
Next Cert	FEB 2022

Tolerance +/-	
0-10000	20
10000-40000	40
40000-50000	60

Comments

2

4

8

1

3

5

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
2000	2000	0	400	0
6000	6000	0	2000	0
10000	10000	0	6000	0
20000	20000	0	10000	0
20020	20020	0	20020	0
30020	30020	0	30020	0
38020	38020	0	38020	0
43600	43600	0	43600	0

Repeatability

Load	Ind	Ind	Ind

Repeatability

Load	Ind	Ind	Ind
38020	38020	38020	38020

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Eccentric Load 10000lb

Reading	Dev	Reading	Dev
1	10000	2	10000
3	10000	4	10000
5	10000	6	10000
7		8	

Signature Name D. Waller Date 3-2-2021
C1.3.88.1M