

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

KENNY WASTE MANAGEMENT
GROBY ROAD NORTH
AUDENSHAM
MANCHESTER

LOCATION **M34 5HT**
SITE

Machine Details

Cust ref	WEIGHBRIDGE
Make	SBA-M
Model	DD700
Ser No	T203900298
Capacity	50/000 / kg
e/d	1 / 20 / kg
Working cap	


Test Weights

Set No.	Cert No.	Date Cal
WB1-22	TR12696	20.1.25

Cert Interval	YEARLY	
Next Cert	MARCH	2026

Tolerance +/-	kg.
0-10000	20
10000-40000	40
40000-50000	60

Comments **Make up weights for = 5200kg**
1orry = 21580kg



AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
10000	10000	0	400	0
18000	18000	0	10000	0
40780	40780	0	40780	0
44780	44780	0	44780	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
40780	40780	40780	40780

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 10000kg

Reading	Dev	Reading	Dev
1 10000	0	2 10000	0
3 10000	0	4 10000	0
5 10000	0	6 10000	0
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name **D. Waller** Date **19.3.2025**
C1.3.88.1N