

Weightron Bilanciai Ltd,
 Titan Works
 Bridge Way off Broombank Road,
 Chesterfield Trading Estate,
 Chesterfield S41 9QJ, UK
 Tel: 01246 260062
 Fax: 01246 260844
 www.weightron.com

Weightron Bilanciai Ltd,
 Unit 4b, Block 4
 Westpark, Chelston
 Wellington TA21 9AD, UK
 Tel: 01823 662355
 Fax: 01823 6621017
 www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the Calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Kenny Waste Management Ltd	
Lester Road	
Worsley Trading Estate	
Little Hulton	
Manchester	
Greater Manchester	
M38 0PT	
Location	Site

Machine Details

Cust ref	Weighbridge	
Make	Bilanciai	
Model	D410 Analogue	
Serial No	T130203731	
Capacity	50000	kg
e/d	20	kg
Working cap	44000	

Test Weights

Set No.	Cert No.	Date Cal
wb 1-22	tr13091	28/02/2027

Cert Interval	12 Months
Next Cert	Feb 2027

Tolerance +/-	kg
0 - 10000	20
10000 - 40000	40
40000 - 50000	60

Comments
flt 5320kg wtu 22980kg

AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
0	0	0	0	0
3000	3000	0	3000	0
6000	6000	0	6000	0
9000	8980	-20	8980	-20
12000	11980	-20	11980	-20
16000	15980	-20	15980	-20
21320	21300	-20	21300	-20
23000	22980	-20	22980	-20
44320	44300	-20	44300	-20

DEFINITIVE LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev

Repeatability

Load kg	Ind	Ind	Ind
44320	44300	44300	44300

Repeatability

Load kg	Ind	Ind	Ind

Eccentric Load 8320

Reading	kg	Dev	Reading	kg	Dev
1	8300	-20	2	8300	-20
3	8320	0	4	8320	0
5	8320	0	6	8320	0
7			8		
9			10		
11			12		

Eccentric Load

Reading	kg	Dev	Reading	kg	Dev
1			2		
3			4		
5			6		
7			8		
9			10		
11			12		

Signature:

Name: Martin Roberts

Date: 27/02/2026 14:28