

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com

Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 6621017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the Calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Kenny Waste Management Ltd - Groby Road	
Groby Road	
Audenshaw	
Manchester	
Greater Manchester	
M34 5HT	
Location	Site Entrance

Machine Details

Cust ref	Weighbridge	
Make	Bilanciai	
Model	DD700 ABS Digital Indicator	
Serial No	T203900298	
Capacity	50000	kg
e/d	20	kg
Working cap	44000	

Test Weights

Set No.	Cert No.	Date Cal
wb 1-22	tr13091	28/02/2027

Cert Interval	12 Months
Next Cert	Feb 2027

Tolerance +/-	kg
0 - 10000	20
10000 - 40000	40
40000 - 50000	60

Comments

flt 5320kg
wtu 23040kg

AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
0	0	0	0	0
3000	3000	0	3000	0
6000	6000	0	6000	0
9000	9000	0	9000	0
12000	12000	0	12000	0
16000	16000	0	16000	0
21320	21320	0	21320	0
23040	23040	0	23040	0
44360	44360	0	44360	0

DEFINITIVE LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev

Repeatability

Load kg	Ind	Ind	Ind
44360	44360	44360	44360

Repeatability

Load kg	Ind	Ind	Ind

Eccentric Load 8320

Reading kg	Dev	Reading kg	Dev
1 8320	0	2 8320	0
3 8320	0	4 8320	0
5 8320	0	6 8320	0
7		8	
9		10	
11		12	

Eccentric Load

Reading kg	Dev	Reading kg	Dev
1		2	
3		4	
5		6	
7		8	
9		10	
11		12	

Signature:

Name: Martin Roberts

Date: 27/02/2026 15:48