

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Kenny Waste Management
LESTER RD
WORSLEY TRADING ESTATE
LOCATION M38 OPT.

Machine Details

Cust ref	WEIGHTBRIDGE
Make	BILANCIAI
Model	D110
Ser No	T130203731
Capacity	50 / 000 / kg
e/d	/ 20 / kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
F001-48	TR12646	13/11/24

Cert Interval	12 MONTHS
Next Cert	JAN 2026

Tolerance +/-	kg
0- 10000	20
10000-40000	40
40000-50000	60

Comments

.....

.....

.....

.....

.....

2	4	6
1	3	5

AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
0	0	0	0	0
2000	2000	0	2000	0
10000	10000	0	10000	0
16000	16020	+20	16020	0
20300	20300	0	20300	0
28000	28000	0	28000	0
38000	38000	0	38000	0
39900	39900	0	39900	0
44200	44200	0	44200	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44200	44200	44200	44200

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 6300 kg

Reading	Dev	Reading	Dev
1 6320	+20	2 6320	+20
3 6280	-20	4 6280	-20
5 6300	0	6 6300	0
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature

Name SHANE TOLIVER

Date 22/1/25